Review 1: "Hitting the diagnostic sweet spot: Point-of-care SARS-CoV-2 salivary antigen testing with an off-the-shelf glucometer"

Valeria Ghisetti¹

¹Director, Public Health, Microbiology and Virology, Amedeo di Savoia Hospital, Italy

Published on: Oct 30, 2020

DOI: 10.1162/2e3983f5.1565f34b

License: [Creative Commons Attribution 4.0 International License \(CC-BY 4.0\)](https://creativecommons.org/licenses/by/4.0/)

RR:C19 Evidence Scale rating by reviewer:

- **Strong.** The main study claims are very well-justified by the data and analytic methods used. There is little room for doubt that the study produced has very similar results and conclusions as compared with the hypothetical ideal study. The study's main claims should be considered conclusive and actionable without reservation.

Review:

Significant barriers to the diagnosis of SARS-CoV-2 infection continue to hamper population-based screening and efforts are required to contain the COVID-19 pandemic in the absence of effective antiviral therapeutics or vaccines. The Authors report an aptamer-based SARS-CoV2 salivary antigen assay employing low-cost reagents (\$3.20/test). The test was quantitative and easy to use with the chance of connecting to smartphones providing an opportunity to integrate with contract tracing apps, medical providers, and electronic medical records. In clinical testing, the developed assay detected SARS CoV-2 infection in patient saliva within one hour, with 100% sensitivity and 100% specificity. This approach can provide an inexpensive, rapid, and accurate diagnostic for screening of SARS-CoV-2 infection in clinical settings as well as in mass-population based screening.