Review of "Soft drinks can be misused to give false "false positive" SARS-CoV-2 lateral flow device results"

Lorenzo Azzi¹

¹Assistant Professor University of Insubria: Università degli Studi dell'Insubria Department of Medicine and Surgery 10, via G. Piatti Varese, VA 21100 ITALY

Published on: Oct 18, 2021

License: [Creative Commons Attribution 4.0 International License \(CC-BY 4.0\)](https://creativecommons.org/licenses/by/4.0/)

To read the original manuscript, click the link above.

Summary of Reviews: This preprint claims that soft drinks can be misused to generate false-positive results on SARS CoV-2 lateral flow assays. Since the study was done on only one type of test kit, the reviewer raised concerns about its generalizability as this may not be the case in other kits.

Reviewer 1 (Lorenzo Azzi) | ☐☐☐☐

RR:C19 Strength of [Evidence Scale Key](#)

☐☐☐☐ = Misleading

☐☐☐☐ = Not Informative

☐☐☐☐ = Potentially Informative

☐☐☐☐☐ = Reliable

☐☐☐☐☐☐ = Strong

To read the reviews, click the links below.